

CERTIFICATE OF CONFORMITY

We herewith declare that the following product meet the provisions of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices as amended by Directive 2007/47/EC and all supporting documentation is retained at the premise of the manufacturer.

| | |
|--|---|
| EC Certificate(s) | G1 11 10 61155 005 |
| Manufacturer Address | TERANG NUSA SDN. BHD. 1, Jalan 8, Pengkalan Chepa 2, Industrial Zone, 16100 Kota Bharu, Kelantan, MALAYSIA |
| European Representative | EMERSON & CO. S.R.L. Piazza Della Vittoria 10/12 16121 Genova ITALY Tel : 0039-010-543 463 Fax: 0039-010-565 480 |
| Brand | MAXITEX UG |
| Product Description | Sterile Latex Surgical Green Gloves Powderfree |
| Classification (MDD 93/42, Annex IX): | IIa (Rule 7) |
| Conformity Assessment | Annex II excluding 4 |
| Standards | ISO 15223-1:2012, EN 455-1:2000, EN 455-2:2009+A2:2013 and EN 455-3:2006 ISO 14971:2009, ISO 10993-1:2009, ISO 10282:2002, ISO 11607-1:2006 and ISO 11137-1:2006 |
| Notified Body | TÜV SÜD Product Service GmbH, Zertifizierstelle, Ridlerstraße 65, 80339 München, Germany (CE 0123) |

Kota Bharu, Malaysia, 14th July 2014


Ranayu MZ
Management Representative

TERANG NUSA Sdn. Bhd. (164501-X)

1, Jalan 8, Pengkalan Chepa 2 Industrial Zone, 16100 Kota Bharu, Malaysia.
Tel : + 609 7747171 Fax : 609 7747757 e-mail : sales@terangnusa.com

www.terangnusa.com