

PAUL HARTMANN AG
Paul-Hartmann-Strasse 12
89522 Heidenheim

Phone: +49 (0) 7321 36-0
Fax: +49 (0) 7321 36-3636
hartmann.info

P.O. Box 1420
89504 Heidenheim
Germany



Going further
for health

PAUL HARTMANN AG, P.O. Box 1420, 89522 Heidenheim, Germany

EC-Declaration of Conformity for Cosmopor I.V. transparent

Heidenheim, 31 January 2020

We herewith declare,

Object of the declaration: Cosmopor I.V. transparent (1448)

which has been first placed on the market by PAUL HARTMANN AG, meets the applicable provisions, especially the Essential Requirements of the following EU-legislation:

- **Council Directive 93/42/EEC for medical devices (MDD)**

The required conformity assessment procedure according to Annex VII in connection with Annex V has been performed and the technical documentation is kept available.

This EC-Declaration of Conformity is issued under the sole responsibility of the PAUL HARTMANN AG.

The sterilization processes are under the supervision of the Notified Body:

TÜV SÜD Product Service GmbH, DE-80339 München, Ridlerstr. 65, Identification No. 0123.

Medical Device: **Class I sterile** acc. to rule 4 (1)
(acc. to Annex IX of the directive)

UMDNS: 10-288

PAUL HARTMANN AG

François Georgelin

Head of Business Division Wound Management

ppa.

Stefan Fischer

Senior Vice President Regulatory Affairs

This document is valid until: 2024-05-25

IILN 040 9500 00000 0

Vorstand/ Management Board: Britta Fünfstück
(Vorstandsvorsitzende/ CEO). François Georgelin.
Dr. Raymund Heinen. Michel Kuehn.

Aufsichtsratsvorsitzender/ Chairman of the Supervisory Board:
Fritz-Jürgen Heckmann

Sitz Heidenheim
Amtsgericht Ulm HRB 661090

Registered Office Heidenheim

Commercial Register of the District Court of Ulm file no. HRB 661090