

EU Declaration of Conformity for Medical Devices Class I

We herewith declare, that the following medical device

HARTMANN Stérilux Coton Pads
Article Number: 918 480/0

of the product group

Group No.	Product Group	Class acc. to MDR (EU) 2017/745	Rule	UMDNS	GMDN
1.01	Medical Wadding	I	1	15-216	58234

which was first placed on the market by CMC Consumer Medical Care GmbH, meets the applicable provisions, especially the general safety and performance requirements of the Regulation (EU) 2017/745 of the European Parliament and of the Council of 5. April 2017 on medical devices.

The Conformity Assessment Procedure according to Article 52 (7) of Regulation (EU) 2017/745 has been performed and the Technical Documentation according to Annexes II and III is kept available.

The EU Declaration of Conformity is issued under the sole responsibility of the CMC Consumer Medical Care GmbH (Registration Number DIMDI DE/CA37/1661).

i.V. 

Dr. Benjamin Wenzel
Director R&D & RA

i.V. 

Susanne Wolpert
Head of R&D & RA

This document is valid until: 2022-06-10

CMC Consumer Medical Care GmbH

Mailing address
Eichendorffstraße 12-14
89567 Sontheim/Brenz, Deutschland

Phone +49 (73 25) 92 44 0
Fax +49 (73 25) 92 44 492

Head Office
Eichendorffstraße 12-14
89567 Sontheim/Brenz, Deutschland

Internet www.cmc-group.de
E-Mail info@cmc-group.de

District Court Ulm HRB 726 009

VA TIN DE 814313953

CEO (Chairman of the Management Board)
Dr. Rainer Mangold

Annex: Applicable Standards

DIN EN	1041		2013	Information supplied by the manufacturer of medical devices.
EN ISO	14971		2012	Application of risk management to medical devices
EN	15223	-1	2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements
DIN EN ISO	13485		2016	Medical devices -- Quality management systems -- Requirements for regulatory purposes
DIN EN	62366	-1	2017	Medical devices - Application of usability engineering to medical devices
ISO	10993	-1	2018	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management system
DIN EN ISO	10993	-5	2009	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity
EN ISO	10993	-10	2020	Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization
EN ISO	10993	-12	2012	Biological evaluation of medical devices - Part 12: Sample preparation and reference materials
EN ISO	10993	-18	2020	Biological evaluation of medical devices - Part 18: Chemical characterization of medical device materials within a risk management process
Ph. Eur. 10.0, 2316	01/2008: 0036		2008	Cotton, absorbent
Ph. Eur. 10.0, 4189	02/2008: 0034		2008	Viscose Wadding, absorbent
Ph. Eur. 10.0, 201	07/2010: 20612		2010	Microbiological evaluation of non-sterile products: Microbial enumeration test
Ph. Eur. 10.0, 205	04/2010: 20613		2010	Microbiological evaluation of non-sterile products: Test for specific micro-organisms
Ph. Eur. 10.0, 628	07/2017: 50106		2017	Alternative methods for microbiological quality control

CMC Consumer Medical Care GmbH

Mailing address
Eichendorffstraße 12-14
89567 Sontheim/Brenz, Deutschland

Phone +49 (73 25) 92 44 0
Fax +49 (73 25) 92 44 492

Head Office
Eichendorffstraße 12-14
89567 Sontheim/Brenz, Deutschland

Internet www.cmc-group.de
E-Mail info@cmc-group.de

District Court Ulm HRB 726 009

VA TIN DE 814313953

CEO (Chairman of the Management Board)
Dr. Rainer Mangold