

PAUL HARTMANN AG  
Paul-Hartmann-Strasse 12  
89522 Heidenheim  
P.O. Box 1420  
89504 Heidenheim  
Germany

Phone: +49 (0) 7321 36-0  
Fax: +49 (0) 7321 36-3636  
**hartmann.info**



## EU Declaration of Conformity

Heidenheim, 2021-09-09

### Object(s) of the declaration:

#### HydroClean mini (2850)

We herewith declare under our sole responsibility that the medical devices listed above, first placed on the market by PAUL HARTMANN AG, comply with the applicable provisions, in particular, the

- General Safety and Performance Requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5. April 2017 on medical devices.

The objects of the declaration have been identified as medical devices in risk class IIb according to classification rule 4 indent 2 in Annex VIII of Regulation (EU) 2017/745. The conformity assessment procedure according to Article 52 (4) and Annex IX has been performed and the Technical Documentation is kept available.

The conformity assessment procedure is under the supervision of the Notified Body:  
TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339 München, Germany, ID-Nr. 0123.

(High-Level) Intended Purpose: Single-use, sterile, non-medicated dressings suitable for the treatment of wounds

Basic UDI-DI: 40495002850L5

SRN: Single Registration Number of Manufacturer: DE-MF-000005861

PAUL HARTMANN AG

ppa.

**François Georgelin**

Head of Business Division

Wound Management

**Stefan Fischer**

Head of Regulatory Affairs

Valid until (yyyy-mm-dd): 2022-03-01

ILN 040 9500 00000 0

Vorstand/Management Board: Britta Fünfstück  
(Vorsitzende des Vorstands/CEO), François Georgelin,  
Stefan Grote, Dr. Raymund Heinen, Stefan Müller  
Aufsichtsratsvorsitzender/Chairman of the Supervisory Board:  
Fritz-Jürgen Heckmann

Sitz Heidenheim  
Amtsgericht Ulm HRB 661090  
Registered Office Heidenheim  
Commercial Register of the District Court of Ulm file no. HRB  
661090