

## Foliodrape® Protect Plus, Laparoscopy Set II reinforced (Art. No. 938 713/2)

Spec.-No.:	D 6.1003a
Department:	CMO-DOE
Date:	2014-06-01

### 1. General Product Description

*Foliodrape® Protect Plus Laparoscopy Set II reinforced* for all kinds of laparoscopic surgeries, e.g. cholecystectomy, herniotomy, appendectomy.

The *Foliodrape® Protect Plus Laparoscopy Set II reinforced* carries the CE-mark according to EU directive 93/42/EWG for medical devices and is classified as class I sterile medical devices. A conformity assessment has been performed for *Foliodrape® Protect Plus Laparoscopy Set II reinforced* and it has been shown to be in compliance with all applicable requirements of the above mentioned directive.

The safe use of *Foliodrape® Protect Plus Laparoscopy Set II reinforced*, therefore, is ensured if the product is used in line with the intended purpose.


### 2. Application / Indication


*Foliodrape® Protect Plus Laparoscopy Set II reinforced* is a disposable product for sterile draping of the patient and the theatre equipment.

### 3. Presentations

<b>Art. No. 938 713/2</b>	<b>Foliodrape® Protect Plus Laparoscopy Set II reinforced</b>
	sterile, 5 Sets per dispenser, 1 dispenser per transport carton
	Dimensions of dispenser: 545 x 374 x 234 mm (L x W x H)
	Dimensions of transport carton: 590 x 390 x 250 mm (L x W x H)

### 4. Product Characteristics

	1 reinforced table cover 140 x 190 cm
	1 mayo stand cover 80 x 145 cm (telescopic folding)
	1 laparoscopy drape 250 x 300 cm with one fenestration (28 x 32 cm), two transparent instrument bags and integrated leggings, reinforced (reinforced zones: 2 x 50 x 15,5 cm)
	4 cellulose towels 30 x 33 cm
<b>Sterilization:</b>	With ethylene oxide gas according to EN 556-1 and EN 11135-1
<b>Self adhesives:</b>	Skin-friendly synthetic rubber, latex free and colophony free

		<h1>Technical Data Sheet</h1>
--	--	-------------------------------

<b>Foliodrape® Protect Plus, Laparoscopy Set II reinforced (Art. No. 938 713/2)</b>		Spec.-No.:	D 6.1003a
		Department:	CMO-DOE
		Date:	2014-06-01

5. Product Requirements

*Foliodrape® Protect Plus* corresponds to all requirements of European standard for surgical drapes EN 13795, for the highest performance level “High performance, critical area”.

*For Details see Technical Data sheet of the base material (D 6.1005a, D 6.1005b), mayo stand cover (D 6.1005c) and back table cover (D 6.1005d).*

6. Labelling

Lot-No. with 9-digit code:

e.g.: 8023XXXXX

LOT	8	023	XXXXX
	year of production	serial production order	for internal use

Expiry date:

e.g. ⌚ 2019 – 01  
year month

Shelf life: 5 years

The product will be sterile during this period as long as the packaging is not damaged and the storage conditions are kept.

7. Packaging

*Foliodrape® Protect Plus Laparoscopy Set II reinforced* is sealed into sterile packaging (peeling pouch) according to EN ISO 11607 and EN 868. The peel pouch is packed into folding box and transport carton, sealed with adhesive tapes and stored on euro-pallet.

		<h1>Technical Data Sheet</h1>
--	--	-------------------------------

<b>Foliodrape® Protect Plus, Laparoscopy Set II reinforced (Art. No. 938 713/2)</b>		Spec.-No.:	D 6.1003a
		Department:	CMO-DOE
		Date:	2014-06-01

Date: 2014-06-01

PAUL HARTMANN AG  
Development OTM & Exam. Gloves (CMO-DOE)

i. A.



Dr. Evelyn Hartmann