
		Technical Data Sheet		
Foliodrape® Protect, Ophthalmology Set (Art. No. 938 935)			Spec.-No.:	D 6.1012g
			Department:	CMO-DOE
			Date:	2015-01-20

#### 1. General Product Description

*Foliodrape® Protect Ophthalmology Set* for all kind of eye surgery.

The *Foliodrape® Protect Ophthalmology Set* carries the CE-mark according to EU directive 93/42/EWG for medical devices and is classified as class I sterile medical devices.

A conformity assessment has been performed for *Foliodrape® Protect* and it has been shown to be in compliance with all applicable requirements of the above mentioned directive.

The safe use of *Foliodrape® Protect Ophthalmology Set*, therefore, is ensured if the product is used in line with the intended purpose.


#### 2. Application / Indication



*Foliodrape® Protect Ophthalmology Set* is a disposable product for sterile draping of the patient and the theatre equipment.

#### 3. Presentations

<b>Art. No. 938 935</b>	<b><i>Foliodrape® Protect Ophthalmology Set</i></b>
	sterile, 9 Sets per dispenser, 1 dispenser per transport carton
	Dimensions of dispenser: 545 x 374 x 234 mm (L x W x H)
	Dimensions of transport carton: 590 x 390 x 250 mm (L x W x H)

#### 4. Product Characteristics

<b>Content:</b>	1 back table cover reinforced 140 x 160 cm
	1 ophthalmology drape 170 x 280 cm with fenestration (6,5 x 4 cm), self-adhesive and surrounding fluid collection bag
<b>Sterilization:</b>	With ethylene oxide gas according to EN 556-1 and EN 11135-1
<b>Self adhesives:</b>	Skin-friendly synthetic rubber, latex free and colophony free

		Technical Data Sheet		
Foliodrape® Protect, Ophthalmology Set (Art. No. 938 935)			Spec.-No.:	D 6.1012g
			Department:	CMO-DOE
			Date:	2015-01-20

## 5. Product Requirements

*Foliodrape® Protect Ophthalmology Set* corresponds to all requirements of European standard for surgical drapes EN 13795, for the highest performance level "High performance, critical area".

*For Details see Technical Data sheet of the base material (D 6.1005a) and back table cover (D 6.1005d).*


## 6. Labelling

Lot-No. with 9-digit code:

e.g.: 4023XXXXX

LOT	4	023	XXXXX
	year of production	serial production order	for internal use

Expiry date:

e.g.  2019 – 01  
year month

Shelf life: 5 years

The product will be sterile during this period as long as the packaging is not damaged and the storage conditions are kept.

## 7. Packaging

*Foliodrape® Protect Ophthalmology Set* is sealed into sterile packaging (peeling pouch) according to EN ISO 11607 and EN 868. The peel pouch is packed into folding box and transport carton, sealed with adhesive tapes and stored on euro-pallet.

Date: 2015-01-20

PAUL HARTMANN AG  
Development OTM & Exam. Gloves (CMO-DOE)

i. A.



Dr. Evelyn Hartmann